

# Annex to the EC Certificate No. 50539-17-05

Valid from 2020-04-29 to 2024-05-26

Revision status of the annex: 0 dated 2020-04-29

Devices/device categories included in the certificate:

## Class I m:

For the products listed below, the review of the Quality System refers exclusively to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

- Non active medical devices with measuring function
  - non invasive aneroid sphygmomanometers

## Class II a:

- Devices for monitoring vital physiological parameters
  - Noninvasive blood pressure units and systems
  - Thermometer, electronic
  - Thermometer, infrared
  - 24-hour ambulatory blood pressure monitor
- Evaluation software for blood pressure units and systems



Ruth Delbeck-Bayer  
DEKRA Certification GmbH, Stuttgart, 2020-04-23  
Notified Body ID-number: 0124

DEKRA Certification GmbH \* Handwerkstraße 15 \* D-70565 Stuttgart \* [www.dekra-certification.de](http://www.dekra-certification.de)